

Ref. No: 2613
Date: 08/06/26
Subject: Halogen lighting systems

REQUEST & RESPONSE

I would be grateful if you could provide the following information regarding the use of halogen lighting within your Trust's estate.

1. Does the Trust currently use halogen lighting systems in any operating theatres (including surgical theatre lighting and/or general theatre illumination)?

Across MWL Operating theatres there are LED or fluorescent lights only

2. Does the Trust currently use halogen light bulbs or halogen lighting systems elsewhere within the hospital estate (including wards, corridors, outpatient areas, offices, plant rooms, or other facilities)?

No Halogen light fittings present on Southport or Ormskirk Hospital sites, across Whiston and St Helen Sites We have 50w/35w halogen light fittings located within medical examination rooms, these are located within A&E, Delivery Suite, Ward 3E and any room listed as an examination room within the hospital.

Once halogen light asset experiences a failure, they are replaced with a like-for-like LED inspection light.

3.If the answer to either question 1 or 2 is yes, please provide:

The date(s) or year(s) when the relevant lighting systems or bulbs were purchased.

2008 - 2010

The date(s) or year(s) when they were installed.

2008 – 2010

2010

4.If the answer to either question 1 or 2 is yes, please provide:

The manufacturer name.

Brandon medical

The model number, product code, or other identifying reference for the lighting

Coolview® C35
Coolview C50 FX/GX/SX
Coolview C35 FX/GX/SX
Coolview C20S3